

CM Connect

Library guide for hybrid abutments,
bridges, bars and prosthesis.

3shape 

Use of CM Connect Abutments Library in CAD dental systems

CM Connect Scan Abutments allow scans on implant level or abutment level. This depends on the type of connection and custom impression or intraoral scan.

The choice of the appropriate library connection depends also on the design of the prosthesis to be milled. CM Connect Library allows restorations on CM Connect Ti base, CM Connect AURUMBase® or restorations to be screwed direct on Straumann® synOcta® abutment, Nobel Biocare Multi- unit® abutment or ASTRA UniAbutment®.

Please note that in case of direct milling the connection is non-engaging only.

Depending on the CAD system, the library structure and connection names differ slightly from each other.

3Shape Dental System – implant level scan

























For 3Shape Dental System on this level the following categories are available:

- CM Connect Ti base – for restorations connected with the implant via the Titanium Ti base. The scanning needs to be performed with the Scan Body for the specific implant connection. Note: This category includes parts scanned from abutment level - Multi-Unit abutment, UniAbutment® and synOcta® abutment - with its specific Scan Abutment (see Table 1.1).
- CM Connect MULTIUNIT & UNIABUTMENT PLUS Ti base – for restorations connected with the transepithelial abutment via the Ti base. The scanning needs to be performed with the Scan Body for the specific implant connection. The desired Multi-Unit abutment or UniAbutment® will be mounted virtually in the CAD Dental System. Please note that in this case the Ti base closing line (between the prosthesis and Ti base chamfer) depends on the proper kit selection (abutment height in mm), (see Table 1.2).
- CM Connect MULTIUNIT & UNIABUTMENT DIRECT – for restorations connected directly with the transepithelial abutment. The scanning needs to be performed with the Scan Body for the specific implant connection. The desired Multi-Unit abutment or UniAbutment® will be mounted virtually in the CAD Dental System. Please note that in this case the closing line between the prosthesis and the transepithelial abutment depends on the proper kit selection (abutment height in mm), (see Table 1.3).
- CM Connect Scan on Uniabutment & Multi-Unit & Synocta (see Table 1.4).
- CM Connect AURUMBase® – for restorations connected with the implant via the AURUMBase®. The scanning needs to be performed with the Scan Body for the specific implant connection. Note: This category includes parts scanned from abutment level - Multi-Unit abutment - with its specific Scan Abutment (see Table 1.5).
- CM Connect Pre-milled Ti – for restorations connected with the implant via the pre-milled abutment made in Titanium. The scanning needs to be performed with the Scan Body for the specific implant connection. Note: The maximum width of the piece is restricted to 10mm.
- CM Connect Pre-milled CoCr – for restorations connected with the implant via the pre-milled abutment made in CoCr. The scanning needs to be performed with the Scan Body for the specific implant connection. Note: The maximum width of the piece is restricted to 14mm.

Take care to select the correct CM Connect Scan Abutment and the corresponding library file, as otherwise the milled prosthesis can differ from desired outcome and errors can occur, for example the level of Ti base settlement.

The tables on the following pages gives an overview of all CM Connect 3Shape Dental System library files.


Table 1.1
CM Connect TI-Base


	Platform	Analogue	Type	Category	System	Kit	Scan Abutments	
Nobel Active®	3.0	41.14.068		CENDRES METAUX TI-BASE	CM NOB ACT TI-BASE	NOB_ACT_3.0 NE 41.15.068	41.52.068 41.53.100	
						NOB_ACT_3.0 NE 41.16.068	41.52.068 41.52.200	
	NP/3.5	41.14.041		CENDRES METAUX TI-BASE	CM NOB ACT TI-BASE	NOB_ACT_NP NE 41.15.041	41.50.041 41.52.041 41.53.100	
						NOB_ACT_NP E 41.16.041	41.50.041 41.52.041 41.52.200	
	RP/4.3	41.14.042		CENDRES METAUX TI-BASE	CM NOB ACT TI-BASE	NOB_ACT_RP NE 41.15.042	41.50.042 41.52.042 41.53.100	
						NOB_ACT_RP E 41.16.042	41.50.042 41.52.042 41.52.200	
	WP/5.5	41.14.069		CENDRES METAUX TI-BASE	CM NOB ACT TI-BASE	NOB_ACT_WP NE 41.15.069	41.52.069 41.53.100	
						NOB_ACT_WP NE 41.16.069	41.52.069 41.52.200	
	Nobel Replace Select™	NP/3.5	41.14.004		CENDRES METAUX TI-BASE	CM NOB RPL TI-BASE	NOB_RPL_NP NE 41.15.004	41.50.004 41.52.004 41.53.100
							NOB_RPL_NP E 41.16.004	41.50.004 41.52.004 41.52.200
		RP/4.3	41.14.005		CENDRES METAUX TI-BASE	CM NOB RPL TI-BASE	NOB_RPL_RP NE 41.15.005	41.50.005 41.52.005 41.53.100
							NOB_RPL_RP E 41.16.005	41.50.005 41.52.005 41.52.200
WP/5.0		41.14.006		CENDRES METAUX TI-BASE	CM NOB RPL TI-BASE	NOB_RPL_WP NE 41.15.006	41.50.006 41.52.006 41.53.100	
						NOB_RPL_WP E 41.16.006	41.50.006 41.52.006 41.52.200	
6.0		41.14.097		CENDRES METAUX TI-BASE	CM NOB RPL TI-BASE	NOB_RPL_NP 6.0 41.15.097	41.52.097 41.53.100	
						NOB_RPL_NP 6.0 41.16.097	41.52.097 41.52.200	
Nobel Brånemark®		NP/3.5	41.14.001		CENDRES METAUX TI-BASE	CM NOB BRK TI-BASE	NOB_BRK_NP NE 41.15.001	41.50.001 41.52.001 41.53.100
							NOB_BRK_NP E 41.16.001	41.50.001 41.52.001 41.52.200
		RP/4.1	41.14.002		CENDRES METAUX TI-BASE	CM NOB BRK TI-BASE	NOB_BRK_RP NE 41.15.002	41.50.002 41.52.002 41.53.100
							NOB_BRK_RP E 41.16.002	41.50.002 41.52.002 41.52.200
	WP/5.0	41.14.003		CENDRES METAUX TI-BASE	CM NOB BRK TI-BASE	NOB_BRK_WP NE 41.15.003	41.50.003 41.52.003	
						NOB_BRK_WP E 41.16.003	41.50.003 41.52.003	























 non-engaging

 non-engaging

 Scan Abutment
for desktop scanning
41.50.OXX


 Scan Abutment
for intra-oral scanning
41.52.OXX


 Scan Abutment
on Ti base
41.53.100 41.52.200

	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
3i Osseotite®	NP/3.4	41.14.011	 	CENDRES METAUX TI-BASE	CM 3i OSS TI-BASE	3I_OSS_NP NE 41.15.011	41.50.011 41.52.011 41.53.100
						3I_OSS_NP E 41.16.011	41.50.011 41.52.011 41.52.200
	RP/4.1	41.14.012	 	CENDRES METAUX TI-BASE	CM 3i OSS TI-BASE	3I_OSS_RP NE 41.15.012	41.50.012 41.52.012 41.53.100
						3I_OSS_NP E 41.16.012	41.50.012 41.52.012 41.52.200
	WP/5.0	41.14.013	 	CENDRES METAUX TI-BASE	CM 3i OSS TI-BASE	3I_OSS_WP NE 41.15.013	41.50.013 41.52.013 41.53.100
						3I_OSS_WP E 41.16.013	41.50.013 41.52.013 41.52.200
3i Certain®	NP/3.4	41.14.014	 	CENDRES METAUX TI-BASE	CM 3i CER TI- BASE	3I_CER_NP NE 41.15.014	41.50.014 41.52.014 41.53.100
						3I_CER_NP E 41.16.014	41.50.014 41.52.014 41.52.200
	RP/4.1	41.14.015	 	CENDRES METAUX TI-BASE	CM 3i CER TI-BASE	3I_CER_RP NE 41.15.015	41.50.015 41.52.015 41.53.100
						3I_CER_RP E 41.16.015	41.50.015 41.52.015 41.52.200
	WP/5.0	41.14.016	 	CENDRES METAUX TI-BASE	CM 3i CER TI-BASE	3I_CER_WP NE 41.15.016	41.50.016 41.52.016 41.53.100
						3I_CER_WP E 41.16.016	41.50.016 41.52.016 41.52.200
Straumann® Soft Tissue Level	NNC/3.5	41.14.070	 	CENDRES METAUX TI-BASE	CM STR STLV TI-BASE	NOB_ACT_WP NE 41.15.069	41.52.069 41.53.100
						NOB_ACT_WP NE 41.16.069	41.52.069 41.52.200
	RN/4.8	41.14.009	 	CENDRES METAUX TI-BASE	CM STR STLV TI-BASE	STR_STLV_RN NE 41.15.046	41.50.009 41.52.009
						STR_STLV_RN E 41.16.009	41.50.009 41.52.009
	RN on synOCTA®	41.14.010	 	CENDRES METAUX TI-BASE	CM STR STLV TI-BASE	STR SYN RN NE 41.15.010	41.50.010 41.52.010
						STR SYN RN E 41.16.010	41.50.010 41.52.010
	WN/6.5	41.14.045	 	CENDRES METAUX TI-BASE	CM STR STLV TI-BASE	STR_STLV_WP NE 41.15.045	41.50.045 41.52.045
						STR_STLV_WP E 41.16.045	41.50.045 41.52.045
	WN on synOCTA®	41.14.046	 	CENDRES METAUX TI-BASE	CM STR STLV TI-BASE	STR SYN WN NE 41.15.048	41.50.015 41.52.015
						STR SYN WN E 41.16.048	41.50.015 41.52.015





















 non-engaging

 non-engaging

 Scan Abutment
for desktop scanning
41.50.OXX

 Scan Abutment
for intra-oral scanning
41.52.OXX


 Scan Abutment
on Ti base
41.53.100 41.52.200


	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
Straumann® Bone Level	NC/3.3	41.14.043		CENDRES METAUX TI-BASE	CM STR BLV TI-BASE	STR_BLV_NC NE 41.15.043	41.50.043 41.52.043 41.53.100
						TR_BLV_NC E 41.16.043	41.50.043 41.52.043 41.52.200
	RC/4.1	41.14.044		CENDRES METAUX TI-BASE	CM STR BLV TI-BASE	STR_BLV_RC NE 41.15.044	41.50.044 41.52.044 41.53.100
						TR_BLV_RC E 41.16.044	41.50.044 41.52.044 41.52.200
	SRA	41.14.098		CENDRES METAUX TI-BASE	CM STR BLV TI-BASE	STR_BLV_SRA 41.15.098	41.52.098
						STR_BLV_SRA 41.16.098	41.52.098
Zimmer Screw-Vent®	NP/3.5	41.14.017		CENDRES METAUX TI-BASE	CM ZIM SCV TI-BASE	ZIM_SCV_NP NE 41.15.017	41.50.017 41.52.017 41.53.100
						ZIM_SCV_NP E 41.16.017	41.50.017 41.52.017 41.52.200
	RP/4.5	41.14.018		CENDRES METAUX TI-BASE	CM ZIM SCV TI-BASE	ZIM_SCV_RP NE 41.15.018	41.50.018 41.52.018 41.53.100
						ZIM_SCV_RP E 41.16.018	41.50.018 41.52.018 41.52.200
	WP/5.7	41.14.019		CENDRES METAUX TI-BASE	CM ZIM SCV TI-BASE	ZIM_SCV_WP NE 41.15.019	41.50.019 41.52.019 41.53.100
						ZIM_SCV_WP E 41.16.019	41.50.019 41.52.019 41.52.200
Astra tech implant system™ EV	EV/3.6	41.14.059		CENDRES METAUX TI-BASE	CM AST EVO TI-BASE	AST_EVO 3.6_NE 41.15.059	41.50.059 41.52.059 41.53.100
						AST_EVO 3.6_E 41.16.059	41.50.059 41.52.059 41.52.200
	EV/4.2	41.14.060		CENDRES METAUX TI-BASE	CM AST EVO TI-BASE	AST_EVO 4.2_NE 41.15.060	41.50.060 41.52.060 41.53.100
						AST_EVO 4.2_E 41.16.060	41.50.060 41.52.060 41.52.200
	EV/4.8	41.14.061		CENDRES METAUX TI-BASE	CM AST EVO TI-BASE	AST_EVO 4.8_NE 41.15.061	41.50.061 41.52.061 41.53.100
						AST_EVO 4.8_E 41.16.061	41.50.061 41.52.061 41.52.200
	Unia- butment®	41.14.058		CENDRES METAUX TI-BASE	CM AST EVO TI-BASE	AST_EVO uniabutment_ NE 41.15.058	41.50.058 41.52.058
						AST_EVO uniabutment_ NE 41.16.058	41.50.058 41.52.058

 non-engaging

 non-engaging

 Scan Abutment
for desktop scanning
41.50.OXX

 Scan Abutment
for intra-oral scanning
41.52.OXX

 Scan Abutment
on Ti base
41.53.100 41.52.200

	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
Astra tech osseospeed™	NP/3.0	41.14.023	●	CENDRES METAUX TI-BASE	CM AST OSS TI-BASE	AST_OSS_NP NE 41.15.023	41.52.023 41.53.100
			⬡			AST_OSS_NP E 41.16.023	41.50.059 41.52.059 41.52.200
	RP/3.5-4.0	41.14.024	●	CENDRES METAUX TI-BASE	CM AST OSS TI-BASE	AST_OSS_RP NE 41.15.024	41.50.024 41.52.024 41.53.100
			⬡			AST_OSS_RP E 41.16.024	41.50.024 41.52.024 41.52.200
	WP/4.5-5.0	41.14.025	●	CENDRES METAUX TI-BASE	CM AST OSS TI-BASE	AST_OSS_WP NE 41.15.025	41.50.025 41.52.025 41.53.100
			⬡			AST_OSS_WP E 41.16.025	41.50.025 41.52.025 41.52.200
Unia-butment® 20°	41.14.036	●	CENDRES METAUX TI-BASE	CM AST OSS TI-BASE	AST_OSS_Unia-butment®_20° NE 41.15.036	41.50.036 41.52.036	
Unia-butment® 45°	41.14.037	●	CENDRES METAUX TI-BASE	CM AST OSS TI-BASE	AST_OSS_Unia-butment®_45° NE 41.15.037	41.50.037 41.52.037	
Friadent XIVE®	NP/3.4	41.14.038	●	CENDRES METAUX TI-BASE	CM DEN XIV TI-BASE	DEN_XIV_NP NE 41.15.038	41.50.038 41.52.038 41.53.100
			⬡			DEN_XIV_NP E 41.16.038	41.50.038 41.52.038 41.52.200
	RP/3.8	41.14.039	●	CENDRES METAUX TI-BASE	CM DEN XIV TI-BASE	DEN_XIV_RP NE 41.15.039	41.50.039 41.52.039 41.53.100
			⬡			DEN_XIV_RP E 41.16.039	41.50.039 41.52.039 41.52.200
	WP/4.5	41.14.040	●	CENDRES METAUX TI-BASE	CM DEN XIV TI-BASE	DEN_XIV_WP NE 41.15.040	41.50.040 41.52.040 41.53.100
			⬡			DEN_XIV_WP E 41.16.040	41.50.040 41.52.040 41.52.200
5.5	41.14.071	●	CENDRES METAUX TI-BASE	CM DEN XIV TI-BASE	DEN_XIV_5.5 NE 41.15.071	41.52.071 41.53.100	
		⬡			DEN_XIV_5.4 E 41.16.071	41.52.071 41.52.200	
Ankylos® C/X	All platforms, height GH 1,5	41.14.048 41.14.049	●	CENDRES METAUX TI-BASE	CM DEN ANK TI-BASE	DEN ANK NE GH1 41.15.050	41.50.049 41.52.049 41.53.100
			⬡			DEN ANK E GH1 41.16.050	41.50.049 41.52.049 41.52.200
	All platforms, height GH 2,5	41.14.048 41.14.049	●	CENDRES METAUX TI-BASE	CM DEN ANK TI-BASE	DEN ANK NE GH2 41.15.250	41.50.049 41.52.049 41.53.100
			⬡			DEN ANK E GH2 41.16.250	41.50.049 41.52.049 41.52.200

● non-engaging


⬡ non-engaging

📏 Scan Abutment for desktop scanning
41.50.0XX

📏 Scan Abutment for intra-oral scanning
41.52.0XX


📏 Scan Abutment on TI base
41.53.100 41.52.200


	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
Osstem® TS	Mini Yellow	41.14.072		CENDRES METAUX TI-BASE	CM OSS TS TI-BASE	OSS_TS MINI_NE 41.15.072	41.50.072 41.52.072 41.53.100
						OSS_TS MINI_E 41.16.072	41.50.072 41.52.072 41.52.200
	STD Regular	41.14.073		CENDRES METAUX TI-BASE	CM OSS TS TI-BASE	OSS_TS STD_NE 41.15.072	41.50.073 41.52.073 41.53.100
						OSS_TS STD_E 41.16.072	41.50.073 41.52.073 41.52.200
Camlog®	3.3	41.14.064		CENDRES METAUX TI-BASE	CM CAMLOG TI-BASE	CAM_3.3 NE 41.15.064	41.52.064 41.53.100
						CAM_3.3 E 41.16.064	41.52.064 41.52.200
	3.8	41.14.065		CENDRES METAUX TI-BASE	CM CAMLOG TI-BASE	CAM_3.8 NE 41.15.265	41.50.065 41.52.065 41.53.100
						CAM_3.8 E 41.16.265	41.50.065 41.52.065 41.52.200
	4.3	41.14.066		CENDRES METAUX TI-BASE	CM CAMLOG TI-BASE	CAM_4.3 NE 41.15.266	41.50.066 41.52.066 41.53.100
						CAM_4.3 E 41.16.266	41.50.066 41.52.066 41.52.200
	5.0	41.14.067		CENDRES METAUX TI-BASE	CM CAMLOG TI-BASE	CAM_5.0 NE 41.15.067	41.52.067 41.53.100
						CAM_5.03 E 41.16.067	41.52.067 41.52.200


















 non-engaging

 non-engaging

 Scan Abutment for desktop scanning
41.50.0XX

 Scan Abutment for intra-oral scanning
41.52.0XX

 Scan Abutment on Ti base
41.53.100 41.52.200

	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
Biohorizns® External	NP/3.7	41.14.020		CENDRES METAUX TI-BASE	CM BH EXT TI- BASE	BH_INTERNAL_NP NE 41.15.051	41.50.020 41.52.020 41.53.100
						BH_INTERNAL_NP E 41.16.051	41.50.020 41.52.020 41.52.200
	RP/4.2	41.14.021		CENDRES METAUX TI-BASE	CM BH EXT TI- BASE	BH_EXTERNAL_RP NE 41.15.052	41.50.021 41.52.021 41.53.100
						BH_EXTERNAL_RP E 41.16.052	41.50.021 41.52.021 41.52.200
	WP/5.2	41.14.022		CENDRES METAUX TI-BASE	CM BH EXT TI- BASE	BH_EXTERNAL_WP NE 41.15.053	41.50.022 41.52.022
						BH_EXTERNAL_WP E 41.16.053	41.50.022 41.52.022
MIS® Seven	NP/3.7	41.14.056		CENDRES METAUX TI-BASE	CM MIS SEV TI-BASE	MIS_SEV_NP_NE 41.15.056	41.50.056 41.52.056 41.53.100
						MIS_SEV_NP_E 41.16.056	41.50.056 41.52.056 41.52.200
	SP/4.2	41.14.095		CENDRES METAUX TI-BASE	CM MIS SEV TI-BASE	MIS_SEV_SP_NE 41.15.095	41.50.017 41.52.017 41.53.100
						MIS_SEV_SP_E 41.16.095	41.50.017 41.52.017 41.52.200
	WP/5.2	41.14.096		CENDRES METAUX TI-BASE	CM MIS SEV TI-BASE	MIS_SEV_WP_NE 41.15.096	41.50.018 41.52.018 41.53.100
						MIS_SEV_WP_E 41.16.096	41.50.018 41.52.018 41.52.200
Megagen AntRidge®		41.14.057		CENDRES METAUX TI-BASE	CM MGN ANY TI-BASE	MGN_ANY_NE 41.15.057	41.50.057 41.53.100
						MGN_ANY_E 41.16.057	41.50.057 41.52.200
Multi-Unit	RP	41.14.007		CENDRES METAUX TI-BASE	CM Multi-Unit TI-BASE	Multi_Unit_RP NE 41.15.007	41.50.007 41.52.007
	RP	41.14.007		CENDRES METAUX TI-BASE	CM Multi-Unit TI-BASE	Multi_Unit_RP NE 41.15.007/8	41.50.007 41.50.007
	WP	41.14.047		CENDRES METAUX TI-BASE	CM Multi-Unit TI-BASE	Multi_Unit_RP E 41.15.054	41.50.047 41.52.047




non-engaging



non-engaging



Scan Abutment
for desktop scanning
41.50.0XX




Scan Abutment
for intra-oral scanning
41.52.0XX






Scan Abutment
on TI base
41.53.100 41.52.200

Table 1.2
CM Connect Multi-Unit &
Uniabutment plus Ti-Base

	Platform	Analogue	Category	System	Kit	Scan Abutments
Multi-Unit on Nobel Active™	NP	41.14.041	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM NOB ACT MULT PLUS INTER	NP 41.15.007+41.40.007/1 (1,5mm) NP 41.15.007+41.40.007/2 (2,5mm) NP 41.15.007+41.40.007/3 (3,5mm)	41.50.041 41.52.041
	RP	41.14.042	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM NOB ACT MULT PLUS INTER	RP 41.15.007+41.40.008/1 (1,5mm) RP 41.15.007+41.40.008/2 (2,5mm) RP 41.15.007+41.40.008/3 (3,5mm) RP 41.15.007+41.40.008/4 (4,5mm)	41.50.042 41.52.042
Multi-Unit on Nobel Replace®	NP	41.14.004	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM NOB RPL MULT PLUS INTER	NP 41.15.007+41.40.004/1 (1mm) NP 41.15.007+41.40.004/2 (2mm) NP 41.15.007+41.40.004/3 (3mm)	41.50.004 41.52.004
	RP	41.14.005	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM NOB RPL MULT PLUS INTER	RP 41.15.007+41.40.005/2 (2mm) RP 41.15.007+41.40.005/3 (3mm) RP 41.15.007+41.40.005/4 (4mm) RP 41.15.007+41.40.005/5 (5mm)	41.50.005 41.52.005
	WP	41.14.063	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM NOB RPL MULT PLUS INTER	WP 41.15.054+41.40.006/1 (1mm) WP 41.15.054+41.40.006/2 (2mm) WP 41.15.054+41.40.006/3 (3mm)	41.50.006 41.52.006
Multi-Unit on Brånemark System®	NP	41.14.001	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM NOB BRK MULT PLUS INTER	NP 41.15.007+41.40.001/1 (1mm) NP 41.15.007+41.40.001/2 (2mm) NP 41.15.007+41.40.001/3 (3mm)	41.50.001 41.52.001
	RP	41.14.002	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM NOB BRK MULT PLUS INTER	RP 41.15.007+41.40.002/2 (2mm) RP 41.15.007+41.40.002/3 (3mm) RP 41.15.007+41.40.002/4 (4mm) RP 41.15.007+41.40.002/5 (5mm)	41.50.002 41.52.002
	WP	41.14.003	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM NOB BRK MULT PLUS INTER	WP 41.15.054+41.40.003/1 (1mm) WP 41.15.054+41.40.003/2 (2mm) WP 41.15.054+41.40.003/3 (3mm)	41.50.003 41.52.003



Platform	Analogue	Category	System	Kit	Scan Abutments	
Multi-Unit on 3i Osseotite®	NP	41.14.011	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM 3i OSS MULT PLUS INTER	NP 41.15.007+41.40.009/1 (1mm)	41.50.011
					NP 41.15.007+41.40.009/2 (2mm)	41.52.011
Multi-Unit on Bone Level™	NP	41.14.043	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM STR BLV MULT PLUS INTER	NP 41.15.007+41.40.043/1 (1,5mm)	41.50.043
					NP 41.15.007+41.40.043/2 (2,5mm)	41.52.043
Multi-Unit on Bone Level™	RP	41.14.044	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM STR BLV MULT PLUS INTER	RP 41.15.007+41.40.044/1 (1,5mm)	41.50.044
					RP 41.15.007+41.40.044/2 (2,5mm)	41.52.044
Multi-Unit on Zimmer Screw-Vent®	NP	41.14.017	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM STR BLV MULT PLUS INTER	NP 41.15.007+41.40.019/1 (1mm)	41.50.017
					NP 41.15.007+41.40.019/2 (2mm)	41.52.017
Multi-Unit on Zimmer Screw-Vent®	RP	41.14.018	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM STR BLV MULT PLUS INTER	RP 41.15.007+41.40.020/1 (1mm)	41.50.018
					RP 41.15.007+41.40.020/2 (2mm)	41.52.018



	Platform	Analogue	Category	System	Kit	Scan Abutments
Uniabutment® on Astra Tech™ Implant System™ EV	RP Uniabutment®	41.14.059	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM AST EVO UNIABUT PLUS INTER	EV3.6 41.15.058+41.40.059/1 (1mm) EV3.6 41.15.058+41.40.059/2 (2mm) EV3.6 41.15.058+41.40.059/3 (3mm) EV3.6 41.15.058+41.40.059/5 (5mm)	41.50.059 41.52.059
	EV 4.2 Uniabutment®	41.14.060	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM AST EVO UNIABUT PLUS INTER	EV4.2 41.15.058+41.40.060/1 (1mm) EV4.2 41.15.058+41.40.060/2 (2mm) EV4.2 41.15.058+41.40.060/3 (3mm)	41.50.060 41.52.060
	EV 4.8 Uniabutment®	41.14.061	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM AST EVO UNIABUT PLUS INTER	EV4.8 41.15.058+41.40.061/1 (1mm) EV4.8 41.15.058+41.40.061/2 (2mm) EV4.8 41.15.058+41.40.061/3 (3mm) EV4.8 41.15.058+41.40.061/5 (5mm)	41.50.061 41.52.061
Uniabutment® on Astra Tech™ Osseospeed®	RP Uniabutment® 20°	41.14.024	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM AST EVO UNIABUT PLUS INTER	RP 20 41.15.036+41.40.013/0 (0,5mm) RP 20 41.15.036+41.40.013/1 (1mm) RP 20 41.15.036+41.40.013/2 (2mm) RP 20 41.15.036+41.40.013/4 (4mm) RP 20 41.15.036+41.40.013/6 (6mm)	41.50.024 41.52.024
	WP Uniabutment® 20°	41.14.025	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM AST EVO UNIABUT PLUS INTER	WP 20 41.15.036+41.40.015/0 (0,5mm) WP 20 41.15.036+41.40.015/1 (1mm) WP 20 41.15.036+41.40.015/2 (2mm) WP 20 41.15.036+41.40.015/4 (4mm) WP 20 41.15.036+41.40.015/6 (6mm)	41.50.025 41.52.025
	RP Uniabutment® 45°	41.14.024	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM AST EVO UNIABUT PLUS INTER	RP 45 41.15.037+41.40.014/0 (0,5mm) RP 45 41.15.037+41.40.014/1 (1mm) RP 45 41.15.037+41.40.014/2 (2mm) RP 45 41.15.037+41.40.014/4 (4mm) RP 45 41.15.037+41.40.014/6 (6mm)	41.50.024 41.52.024
	WP Uniabutment® 45°	41.14.025	CENDRES METAUX MULTIUNIT & UNIA-BUTMENT PLUS TI-BASE	CM AST EVO UNIABUT PLUS INTER	WP 45 41.15.037+41.40.016/0 (0,5mm) WP 45 41.15.037+41.40.016/1 (1mm) WP 45 41.15.037+41.40.016/2 (2mm) WP 45 41.15.037+41.40.016/4 (4mm) WP 45 41.15.037+41.40.016/6 (6mm)	41.50.025 41.52.025








Table 1.3
CM Connect Multi-Unit &
Uniabutment direct

	Platform	Analogue	Category	System	Kit	Scan Abutments
Multi-Unit on Nobel Active™	NP	41.14.041	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM NOB ACT Multi-Unit DIRECT	NP 41.40.007/1 (1,5mm) NP 41.40.007/2 (2,5mm) NP 41.40.007/3 (3,5mm)	41.50.041 41.52.041
	RP	41.14.042	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM NOB ACT Multi-Unit DIRECT	RP 41.40.008/1 (1,5mm) RP 41.40.008/2 (2,5mm) RP 41.40.008/3 (3,5mm) RP 41.40.008/4 (4,5mm)	41.50.042 41.52.042
Multi-Unit on Nobel Replace®	NP	41.14.004	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM NOB RPL Multi-Unit DIRECT	NP 41.40.004/1 (1mm) NP 41.40.004/2 (2mm) NP 41.40.004/3 (3mm)	41.50.004 41.52.004
	RP	41.14.005	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM NOB RPL Multi-Unit DIRECT	RP 41.40.005/2 (2mm) RP 41.40.005/3 (3mm) RP 41.40.005/4 (4mm) RP 41.40.005/5 (5mm)	41.50.005 41.52.005
	WP	41.14.063	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM NOB RPL Multi-Unit DIRECT	WP 41.40.006/1 (1mm) WP 41.40.006/2 (2mm) WP 41.40.006/3 (3mm)	41.50.006 41.52.006
Multi-Unit on Brånemark System®	NP	41.14.001	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM NOB BRK Multi-Unit DIRECT	NP 41.40.001/1 (1mm) NP 41.40.001/2 (2mm) NP 41.40.001/3 (3mm)	41.50.001 41.52.001
	RP	41.14.002	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM NOB BRK Multi-Unit DIRECT	RP 41.40.002/2 (2mm) RP 41.40.002/3 (3mm) RP 41.40.002/4 (4mm) RP 41.40.002/5 (5mm)	41.50.002 41.52.002
	WP	41.14.003	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM NOB BRK Multi-Unit DIRECT	WP 41.40.003/1 (1mm) WP 41.40.003/2 (2mm) WP 41.40.003/3 (3mm)	41.50.003 41.52.003
Multi-Unit on 3i Osseotite®	NP	41.14.011	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM 3i OSS Multi-Unit DIRECT	NP 41.40.009/1 (1mm) NP 41.40.009/2 (2mm) NP 41.40.009/3 (3mm)	41.50.011 41.52.011



	Platform	Analogue	Category	System	Kit	Scan Abutments
Multi-Unit on Bone Level™	NP	41.14.043	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM STR BLV Multi-Unit DIRECT	NP 41.40.043/1 (1,5mm) NP 41.40.043/2 (2,5mm) NP 41.40.043/3 (3,5mm)	41.50.043 41.52.043
	RP	41.14.044	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM STR BLV Multi-Unit DIRECT	RP 41.40.044/1 (1,5mm) RP 41.40.044/2 (2,5mm) RP 41.40.044/3 (3,5mm) RP 41.40.044/4 (4,5mm)	41.50.044 41.52.044
Multi-Unit on Zimmer Screw-Vent®	NP	41.14.017	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM STR BLV Multi-Unit DIRECT	NP 41.40.019/1 (1mm) NP 41.40.019/2 (2mm) NP 41.40.019/3 (3mm) NP 41.40.019/4 (4mm)	41.50.017 41.52.017
	RP	41.14.018	CENDRES METAUX MULTIUNIT & UNI- ABUTMENT DIRECT	CM STR BLV Multi-Unit DIRECT	RP 41.40.020/1 (1mm) RP 41.40.020/2 (2mm) RP 41.40.020/3 (3mm)	41.50.018 41.52.018



	Platform	Analogue	Category	System	Kit	Scan Abutments
Uniabutment® on Astra Tech™ Implant System™ EV	RP Uni-abutment®	41.14.059	CENDRES METAUX MULTIUNIT & UNI-ABUTMENT DIRECT	CM AST EVO UNIABUT PLUS INTER	EV3.6 41.40.059/1 (1mm)	41.50.059
					EV3.6 41.40.059/2 (2mm)	41.52.059
	EV 4.2 Uni-abutment®	41.14.060	CENDRES METAUX MULTIUNIT & UNI-ABUTMENT DIRECT	CM AST EVO UNIABUT PLUS INTER	EV4.2 41.40.060/1 (1mm)	41.50.060
					EV4.2 41.40.060/2 (2mm)	41.52.060
	EV 4.8 Uni-abutment®	41.14.061	CENDRES METAUX MULTIUNIT & UNI-ABUTMENT DIRECT	CM AST EVO UNIABUT PLUS INTER	EV4.8 41.40.061/1 (1mm)	41.50.061
					EV4.8 41.40.061/2 (2mm)	41.52.061
					EV4.8 41.40.061/3 (3mm)	
					EV4.8 41.40.061/5 (5mm)	
Uniabutment® on Astra Tech™ Osseospeed®	RP Uni-abutment® 20°	41.14.024	CENDRES METAUX MULTIUNIT & UNI-ABUTMENT DIRECT	CM AST EVO UNIABUT PLUS INTER	RP 20 41.40.013/0 (0,5mm)	41.50.024
					RP 20 41.40.013/1 (1mm)	41.52.024
	WP Uni-abutment® 20°	41.14.025	CENDRES METAUX MULTIUNIT & UNI-ABUTMENT DIRECT	CM AST EVO UNIABUT PLUS INTER	WP 20 41.40.015/0 (0,5mm)	41.50.025
					WP 20 41.40.015/1 (1mm)	41.52.025
	RP Uni-abutment® 45°	41.14.024	CENDRES METAUX MULTIUNIT & UNI-ABUTMENT DIRECT	CM AST EVO UNIABUT PLUS INTER	RP 45 41.40.014/0 (0,5mm)	41.50.024
					RP 45 41.40.014/1 (1mm)	41.52.024
	WP Uni-abutment® 45°	41.14.025	CENDRES METAUX MULTIUNIT & UNI-ABUTMENT DIRECT	CM AST EVO UNIABUT PLUS INTER	WP 45 41.40.016/0 (0,5mm)	41.50.025
					WP 45 41.40.016/1 (1mm)	41.52.025
					WP 45 41.40.016/2 (2mm)	
					WP 45 41.40.016/4 (4mm)	
					WP 45 41.40.016/6 (6mm)	





Table 1.4
CM Connect scan on Uniabutment &
Multi-Unit & Synocta

	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
Uniabutment® on Astra Tech™ Osseospeed™	Uniabutment® 20°	41.14.036	●	CM UNIABUTMENT DIRECT (Uniabutment Scan Abutment)	CM AST UNIABUTMENT DIRECT (Uniabutment Scan Abutment)	20	41.50.036 41.52.036
	Uniabutment® 45°	41.14.037	●	CM UNIABUTMENT DIRECT (Uniabutment Scan Abutment)	CM AST UNIABUTMENT DIRECT (Uniabutment Scan Abutment)	45	41.50.037 41.52.037
Astra Tech™ Implant System™ EV	Uniabutment®	41.14.058	●	CM UNIABUTMENT DIRECT (Uniabutment Scan Abutment)	CM AST UNIABUTMENT DIRECT (Uniabutment Scan Abutment)	EV Uniabutment	41.50.058 41.52.058
Multi-Unit	RP	41.14.007	●	CM MULTIUNIT DIRECT (Multi-Unit Scan Abutment)	CM AST UNIABUTMENT DIRECT (Uniabutment Scan Abutment)	Multi-Unit	41.50.036 41.52.036
	WP	41.14.047	●	CM MULTIUNIT DIRECT (Multi-Unit Scan Abutment)	CM AST UNIABUTMENT DIRECT (Uniabutment Scan Abutment)	Multi-Unit WP	41.50.037 41.52.037
Straumann® Soft Tissue level witch synOcta® Abutment	RN on synOcta®	41.14.010	●	CM STR SYN DIRECT (Bars and bridges)	CM STR SYN DIRECT (Bars and bridges)	STR SYN RN 51.010	41.50.010 41.52.010
	WN on synOcta®	41.14.046	●	CM STR SYN DIRECT (Bars and bridges)	CM STR SYN DIRECT (Bars and bridges)	STR SYN WN 51.048	41.50.046 41.52.037




non-engaging



non-engaging

























Scan Abutment
for desktop scanning
41.50.OXX



Scan Abutment
for intra-oral scanning
41.52.OXX

Table 1.5
CM Connect AURUMBase®

	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
Nobel Active®	NP	41.14.041		CENDRES METAUX AURUM- Base	CM NOB ACT AURUMBase	NOB_ACT_NP NE 41.36.041	41.50.041 41.52.041 41.53.100
						NOB_ACT_NP E 41.37.041	41.50.041 41.52.041 41.52.200
	RP	41.14.042		CENDRES METAUX AURUM- Base	CM NOB ACT AURUMBase	NOB_ACT_RP NE 41.36.042	41.50.042 41.52.042 41.53.100
						NOB_ACT_RP E 41.37.042	41.50.042 41.52.042 41.52.200
Nobel Replace Select™	NP	41.14.004		CENDRES METAUX AURUM- Base	CM NOB RPL AURUMBase	NOB_RPL_NP NE 41.36.004	41.50.004 41.52.004 41.53.100
						NOB_RPL_NP E 41.37.004	41.50.004 41.52.004 41.52.200
	RP	41.14.005		CENDRES METAUX AURUM- Base	CM NOB RPL AURUMBase	NOB_RPL_RP NE 41.36.005	41.50.005 41.52.005 41.53.100
						NOB_RPL_RP E 41.37.005	41.50.005 41.52.005 41.52.200
	WP	41.14.006		CENDRES METAUX AURUM- Base	CM NOB RPL AURUMBase	NOB_RPL_WP NE 41.36.006	41.50.006 41.52.006 41.53.100
						NOB_RPL_WP E 41.37.006	41.50.006 41.52.006 41.52.200
Nobel Brånemark®	NP	41.14.001		CENDRES METAUX AURUM- Base	CM NOB BRK AURUMBase	NOB_BRK_NP NE 41.36.001	41.50.001 41.52.001 41.53.100
						NOB_BRK_NP E 41.37.001	41.50.001 41.52.001 41.52.200
	RP	41.14.002		CENDRES METAUX AURUM- Base	CM NOB BRK AURUMBase	NOB_BRK_RP NE 41.36.002	41.50.002 41.52.002 41.53.100
						NOB_BRK_RP E 41.37.002	41.50.002 41.52.002 41.52.200
	WP	41.14.003		CENDRES METAUX AURUM- Base	CM NOB BRK AURUMBase	NOB_BRK_WP NE 41.36.003	41.50.003 41.52.003 41.52.003
						NOB_BRK_WP E 41.37.003	41.50.003 41.52.003 41.52.003
3i Osseotite®	NP/3.4	41.14.011		CENDRES METAUX AURUM- Base	CM 3i OSS AURUMBase	3i_OSS_NP NE 41.36.011	41.50.011 41.52.011 41.53.100
						3i_OSS_NP E 41.37.011	41.50.011 41.52.011 41.52.200
	RP/4.1	41.14.012		CENDRES METAUX AURUM- Base	CM 3i OSS AURUMBase	3i_OSS_RP NE 41.36.012	41.50.012 41.52.012 41.53.100
						3i_OSS_RP E 41.37.012	41.50.012 41.52.012 41.52.200
	WP/5.0	41.14.013		CENDRES METAUX AURUM- Base	CM 3i OSS AURUMBase	3i_OSS_WP NE 41.36.013	41.50.013 41.52.013 41.53.100
						3i_OSS_WP E 41.37.013	41.50.013 41.52.013 41.52.200




non-engaging























non-engaging



Scan Abutment
for desktop scanning
41.50.0XX





Scan Abutment
for intra-oral scanning
41.52.0XX

	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
3i Certain®	NP/3.4	41.14.014		CENDRES METAUX AURUM- Base	CM 3I CER AURUMBase	3I_CER_NP NE 41.36.014	41.50.014 41.52.014 41.53.100
						3I_CER_NP E 41.37.014	41.50.014 41.52.014 41.52.200
	RP/4.1	41.14.015		CENDRES METAUX AURUM- Base	CM 3I CER AURUMBase	3I_CER_RP NE 41.36.015	41.50.015 41.52.015 41.53.100
						3I_CER_RP E 41.37.015	41.50.015 41.52.015 41.52.200
	WP/5.0	41.14.016		CENDRES METAUX AURUM- Base	CM 3I CER AURUMBase	3I_CER_WP NE 41.36.016	41.50.016 41.52.016 41.53.100
						3I_CER_NP E 41.37.016	41.50.016 41.52.016 41.52.200
Straumann® Soft Tissue Level	RN/4.8	41.14.009		CENDRES METAUX AURUM- Base	CM STR STLV AURUMBase	STR_STLV_RN NE 41.36.046	41.50.009 41.52.009 41.53.100
						STR_STLV_RN E 41.37.009	41.50.009 41.52.009 41.52.200
	WN/6.5	41.14.045		CENDRES METAUX AURUM- Base	CM STR STLV AURUMBase	STR_STLV_WP NE 41.36.045	41.50.045 41.52.045 41.53.100
						STR_STLV_WP E 41.37.045	41.50.045 41.52.045 41.52.200
Straumann® Bone Level	NC/3.3	41.14.043		CENDRES METAUX AURUM- Base	CM STR BLV AURUMBase	STR_BLV_NC NE 41.36.043	41.50.043 41.52.043 41.53.100
						TR_BLV_NC E 41.37.043	41.50.043 41.52.043 41.52.200
	RC/4.1	41.14.044		CENDRES METAUX AURUM- Base	CM STR BLV AURUMBase	STR_BLV_RC NE 41.36.044	41.50.044 41.52.044 41.53.100
						TR_BLV_RC E 41.37.044	41.50.044 41.52.044 41.52.200
Zimmer Screw-Vent®	NP/3.5	41.14.017		CENDRES METAUX AURUM- Base	CM ZIM SCV AURUMBase	ZIM_SCV_NP NE 41.36.017	41.50.017 41.52.017 41.53.100
						ZIM_SCV_NP E 41.37.017	41.50.017 41.52.017 41.52.200
	RP/4.5	41.14.018		CENDRES METAUX AURUM- Base	CM ZIM SCV AURUMBase	ZIM_SCV_RP NE 41.36.018	41.50.018 41.52.018 41.53.100
						ZIM_SCV_RP E 41.37.018	41.50.018 41.52.018 41.52.200
	WP/5.7	41.14.019		CENDRES METAUX AURUM- Base	CM ZIM SCV AURUMBase	ZIM_SCV_WP NE 41.36.019	41.50.019 41.52.019 41.53.100
						ZIM_SCV_WP E 41.37.019	41.50.019 41.52.019 41.52.200

















 non-engaging

 non-engaging

 Scan Abutment
for desktop scanning
41.50.OXX


 Scan Abutment
for intra-oral scanning
41.52.OXX


 Scan Abutment
on TI base
41.53.100 41.52.200


	Platform	Analogue	Type	Category	System	Kit	Scan Abutments
Astra tech implant system™ EV	EV/3.6	41.14.059		CENDRES METAUX AURUM- Base	CM AST EVO AURUMBase	AST_EVO 3.6_NE 41.36.059	41.50.059 41.52.059 41.53.100
						AST_EVO 3.6_E 41.37.059	41.50.059 41.52.059 41.52.200
	EV/4.2	41.14.060		CENDRES METAUX AURUM- Base	CM AST EVO AURUMBase	AST_EVO 4.2_NE 41.36.060	41.50.060 41.52.060 41.53.100
						AST_EVO 4.2_E 41.37.060	41.50.060 41.52.060 41.52.200
	EV/4.8	41.14.061		CENDRES METAUX AURUM- Base	CM AST EVO AURUMBase	AST_EVO 4.8_NE 41.36.061	41.50.061 41.52.061 41.53.100
						AST_EVO 4.8_E 41.37.061	41.50.061 41.52.061 41.52.200
Astra tech osseospeed™	RP	41.14.024		CENDRES METAUX AURUM- Base	CM AST OSS AURUMBase	AST_OSS_RP NE 41.36.024	41.50.024 41.52.024 41.53.100
						AST_OSS_RP E 41.37.024	41.50.024 41.52.024 41.52.200
	WP	41.14.025		CENDRES METAUX AURUM- Base	CM AST OSS AURUMBase	AST_OSS_WP NE 41.36.025	41.50.025 41.52.025 41.53.100
						AST_OSS_WP E 41.37.025	41.50.025 41.52.025 41.52.200
Friadent XIVE®	NP	41.14.038		CENDRES METAUX AURUM- Base	CM DEN XIV AURUMBase	DEN_XIV_NP NE 41.36.038	41.50.038 41.52.038 41.53.100
						DEN_XIV_NP E 41.37.038	41.50.038 41.52.038 41.52.200
	RP	41.14.039		CENDRES METAUX AURUM- Base	CM DEN XIV AURUMBase	DEN_XIV_RP NE 41.36.039	41.50.039 41.52.039 41.53.100
						DEN_XIV_RP E 41.37.039	41.50.039 41.52.039 41.52.200
	WP	41.14.040		CENDRES METAUX AURUM- Base	CM DEN XIV AURUMBase	DEN_XIV_WP NE 41.36.040	41.50.040 41.52.040 41.53.100
						DEN_XIV_WP E 41.37.040	41.50.040 41.52.040 41.52.200







 non-engaging

 non-engaging


 Scan Abutment
for desktop scanning
41.50.OXX

 Scan Abutment
for intra-oral scanning
41.52.OXX


 Scan Abutment
on Ti base
41.53.100 41.52.200

	Platform	Analogue	Type	Category	System	Kit	Scan Abutments	
Camlog®	3.8	14.065		CENDRES METAUX AURUM- Base	CENDRES METAUX CAMLOG AURUMBase	CAM_3.8 NE 36.065	50.065	52.065
			CAM_3.8 E 36.065			50.065	52.065	
	4.3	14.066		CENDRES METAUX AURUM- Base	CENDRES METAUX CAMLOG AURUMBase	CAM_4.3 NE 37.066	50.066	52.066
			CAM_4.3 E 37.066			50.066	52.066	
Mis® Seven	SP	14.095		CENDRES METAUX AURUM- Base	CENDRES METAUX MIS SEV AURUM- Base	MIS_SEV_SP NE 36.017	50.017	52.017
			MIS_SEV_SP E 37.017			50.017	52.017	
	WP	14.096		CENDRES METAUX AURUM- Base	CENDRES METAUX MIS SEV AURUM- Base	STR_BLV_RC WP NE 36.018	50.018	52.018
			STR_BLV_RC WP E 37.018			50.018	52.018	
Multi-Unit	RP	14.007		CENDRES METAUX AURUM- Base	CENDRES METAUX Multi- UnitAURUM- Base	Multi_Unit_RP NE 36.007	50.007	52.007
	WP	14.047		CENDRES METAUX AURUM- Base	CENDRES METAUX Multi- UnitAURUM- Base	Multi_Unit_RP E 37.054	50.047	52.047

 non-engaging

 non-engaging

 Scan Abutment
for desktop scanning
41.50.0XX

 Scan Abutment
for intra-oral scanning
41.52.0XX




 Scan Abutment
on Ti base
41.53.100 41.52.200

Table 1.6
CM Connect Pre-milled
Abutment Ti


	Platform	Analogue	Category	System	Kit	Scan Abutments	
Nobel Active®	3.0	41.14.068	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB ACT PRE-MILLED TI	NOB_ACT_3.0 E 41.61.068	41.52.068	41.53.100
	NP	41.14.041	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB ACT PRE-MILLED TI	NOB_ACT_NP E 41.61.041ø10mm 41.62.041ø14mm	41.50.041	41.52.041 41.53.100
	RP	41.14.042	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB ACT PRE-MILLED TI	NOB_ACT_RP E 41.61.042ø10mm 41.62.042ø14mm	41.50.042	41.52.042 41.53.100
	WP	41.14.069	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB ACT PRE-MILLED TI	NOB_ACT_WP E 41.61.069	41.52.069	41.53.100
Nobel Replace Select™	NP	41.14.004	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB RPL PRE-MILLED TI	NOB_RPL_NP E 41.61.004	41.50.004	41.52.004 41.53.100
	RP	41.14.005	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB RPL PRE-MILLED TI	NOB_RPL_RP E 41.61.005	41.50.005	41.52.005 41.53.100
	WP	41.14.006	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB RPL PRE-MILLED TI	NOB_RPL_WP E 41.61.006	41.50.006	41.52.006 41.53.100
	6.0	41.14.097	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB RPL PRE-MILLED TI	NOB_RPL_NP 6.0 E 41.61.097	41.52.097	41.53.100
Nobel Brånemark®	NP	41.14.001	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB BRK PRE-MILLED TI	NOB_BRK_NP E 41.61.001	41.50.001	41.52.001 41.53.100
	RP	41.14.002	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB BRK PRE-MILLED TI	NOB_BRK_RP E 41.61.002	41.50.002	41.52.002 41.53.100
	WP	41.14.003	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM NOB BRK PRE-MILLED TI	NOB_BRK_WP E 41.61.003	41.50.003	41.52.003



3i Osseotite®	NP	41.14.011	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM 3i OSS PRE-MILLED TI	3i_OSS_NP E 41.61.011	41.50.011 41.52.011 41.53.100
	RP	41.14.012	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM 3i OSS PRE-MILLED TI	3i_OSS_RP E 41.61.012	41.50.012 41.52.012 41.53.100
	WP	41.14.013	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM 3i OSS PRE-MILLED TI	3i_OSS_WP E 41.61.013	41.50.013 41.52.013 41.53.100
	Platform	Analogue	Category	System	Kit	Scan Abutments
3i Certain®	NP	41.14.014	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM 3i CER PRE-MILLED TI	3i_CER_NP E 41.61.014	41.50.014 41.52.014 41.53.100
	RP	41.14.015	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM 3i CER PRE-MILLED TI	3i_CER_RP E 41.61.015	41.50.015 41.52.015 41.53.100
	WP	41.14.016	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM 3i CER PRE-MILLED TI	3i_CER_WP E 41.61.016	41.50.016 41.52.016 41.53.100
Straumann® Soft Tissue Level	NNC	41.14.070	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM STR STL PRE-MILLED TI	NOB_ACT_WP E 41.61.070	41.52.069 41.53.100
	RN	41.14.009	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM STR STL PRE-MILLED TI	STR STL RN E 41.61.009 ø10mm 41.62.009 ø14mm	41.50.009 41.52.009
	WN	41.14.045	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM STR STL PRE-MILLED TI	STR STL WN E 41.61.045 ø10mm 41.62.045 ø14mm	41.50.045 41.52.045




non-engaging




non-engaging



Scan Abutment
for desktop scanning
41.50.0XX




Scan Abutment
for intra-oral scanning
41.52.0XX



Scan Abutment
on Ti base
41.53.100 41.52.200

	Platform	Analogue	Category	System	Kit	Scan Abutments
Straumann® Bone Level	NC	41.14.043	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM STR BLV PRE-MILLED TI	STR_BLV_NC E 41.61.043 41.62.043	41.50.043 41.52.043 41.53.100
	RC	41.14.044	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM STR BLV PRE-MILLED TI	STR_BLV_RC E 41.61.044 41.62.044	41.50.044 41.52.044 41.53.100
Zimmer Screw-Vent®	NP	41.14.017	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM ZIM SCV PRE-MILLED TI	ZIM_SCV_NP E 41.61.017 41.62.017	41.50.017 41.52.017 41.53.100
	RP	41.14.018	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM ZIM SCV PRE-MILLED TI	ZIM_SCV_RP E 41.61.018	41.50.018 41.52.018 41.53.100
	WP	41.14.019	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM ZIM SCV PRE-MILLED TI	ZIM_SCV_WP E 41.61.019	41.50.019 41.52.019 41.53.100
Astra tech implant system™ EV	EV/3.6	41.14.059	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM AST EVO PRE-MILLED TI	AST_EVO 3.6_E 41.61.059	41.50.059 41.52.059 41.53.100
	EV/4.2	41.14.060	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM AST EVO PRE-MILLED TI	AST_EVO 4.2_E 41.61.060	41.50.060 41.52.060 41.53.100
	EV/4.8	41.14.061	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM AST EVO PRE-MILLED TI	AST_EVO 4.8_E 41.61.061	41.50.061 41.52.061 41.53.100
Astra tech implant Osseospeed™	RP	41.14.024	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM AST OSS PRE-MILLED TI	AST_OSS_RP E 41.61.024	41.50.024 41.52.024 41.53.100
	WP	41.14.025	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM AST OSS PRE-MILLED TI	AST_OSS_WP E 41.61.025	41.50.025 41.52.025 41.53.100
Friadent XIVE®	NP	41.14.038	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM DEN XIV PRE-MILLED TI	DEN_XIV_NP E 41.61.038	41.50.038 41.52.038 41.53.100
	RP	41.14.039	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM DEN XIV PRE-MILLED TI	DEN_XIV_RP E 41.61.039	41.50.039 41.52.039 41.53.100
	WP	41.14.040	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM DEN XIV PRE-MILLED TI	DEN_XIV_WP E 41.61.040	41.50.040 41.52.040 41.53.100
Ankylos® C/X	All platforms, height GH 2,5	41.14.048 41.14.049	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM DEN ANK PRE-MILLED TI	DEN ANK E GH2 41.61.250	41.50.049 41.52.049 41.53.100
	Osstem® TS	Mini Yellow	41.14.072	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM OSS TS PRE-MILLED TI	OSS_TS MINI_E 41.61.072
	STD Regular	41.14.073	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM OSS TS PRE-MILLED TI	OSS_TS STD_E 41.61.072	41.50.073 41.52.073 41.53.100



	Platform	Analogue	Category	System	Kit	Scan Abutments
Camlog®	3.3	41.14.064	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM CAMLOG PRE-MILLED TI	CAM_3.3 E 41.61.064	41.52.064 41.53.100
	3.8	41.14.065	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM CAMLOG PRE-MILLED TI	CAM_3.8 E 41.61.065	41.50.065 41.52.065 41.53.100
	4.3	41.14.066	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM CAMLOG PRE-MILLED TI	CAM_4.3 E 41.61.066	41.50.066 41.52.066 41.53.100
	5.0	41.14.067	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM CAMLOG PRE-MILLED TI	CAM_5.0 E 41.61.067	41.52.067 41.53.100
Biohorizns® External	RP/4.2	41.14.021	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM BH EXT PRE-MILLED TI	BH_EXTERNAL_RP E 41.61.021	41.50.021 41.52.021 41.53.100
	WP/5.2	41.14.022	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM BH EXT PRE-MILLED TI	BH_EXTERNAL_WP E 41.61.022	41.50.022 41.52.022
MIS® Seven	SP	41.14.095	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM MIS SEV PRE-MILLED TI	MIS_SEV_SP_E 41.61.017	41.50.017 41.52.017 41.53.100
	WP	41.14.096	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM MIS SEV PRE-MILLED TI	MIS_SEV_WP_E 41.61.018	41.50.018 41.52.018 41.53.100
Megagen AntRidge®		41.14.057	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM MGN ANY PRE-MILLED TI	MGN_ANY E 41.61.057	41.50.057 41.53.100
Dentium Superline™ & Implantium	RP	41.14.090	CENDRES METAUX PRE-MILLED ABUTMENT TI	CM DENTIUM PRE-MILLED TI	DENTIUM E 41.61.090	41.52.090 41.53.100




non-engaging



non-engaging



Scan Abutment
for desktop scanning
41.50.OXX





Scan Abutment
for intra-oral scanning
41.52.OXX





Scan Abutment
on Ti base
41.53.100 41.52.200

Table 1.7
CM Connect Pre-milled
Abutment CoCr





	Platform	Analogue	Category	System	Kit	Scan Abutments
Nobel Active®	3.0	41.14.068	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB ACT PRE-MILLED CoCr	NOB_ACT_3.0 E 41.71.068	41.52.068 41.53.100
	NP	41.14.041	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB ACT PRE-MILLED CoCr	NOB_ACT_NP E 41.71.041	41.50.041 41.52.041 41.53.100
	RP	41.14.042	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB ACT PRE-MILLED CoCr	NOB_ACT_RP E 41.71.042	41.50.042 41.52.042 41.53.100
	WP	41.14.069	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB ACT PRE-MILLED CoCr	NOB_ACT_WP E 41.71.069	41.52.069 41.53.100
Nobel Replace Select™	NP	41.14.004	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB RPL PRE-MILLED CoCr	NOB_RPL_NP E 41.71.004	41.50.004 41.52.004 41.53.100
	RP	41.14.005	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB RPL PRE-MILLED CoCr	NOB_RPL_RP E 41.71.005	41.50.005 41.52.005 41.53.100
	WP	41.14.006	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB RPL PRE-MILLED CoCr	NOB_RPL_WP E 41.71.006	41.50.006 41.52.006 41.53.100
	6.0	41.14.097	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB RPL PRE-MILLED CoCr	NOB_RPL_NP 6.0 E 41.71.097	41.52.097 41.53.100







Nobel Brånemark®	NP	41.14.001	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB BRK PRE-MILLED CoCr	NOB_BRK_NP E 41.71.001	41.50.001 41.52.001 41.53.100
	RP	41.14.002	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB BRK PRE-MILLED CoCr	NOB_BRK_RP E 41.71.002	41.50.002 41.52.002 41.53.100
	WP	41.14.003	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM NOB BRK PRE-MILLED CoCr	NOB_BRK_WP E 41.71.003	41.50.003 41.52.003
	Platform	Analogue	Category	System	Kit	Scan Abutments
3i Osseotite®	NP	41.14.011	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM 3I OSS PRE-MILLED CoCr	3I_OSS_NP E 41.71.011	41.50.011 41.52.011 41.53.100
	RP	41.14.012	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM 3I OSS PRE-MILLED CoCr	3I_OSS_RP E 41.71.012	41.50.012 41.52.012 41.53.100
	WP	41.14.013	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM 3I OSS PRE-MILLED CoCr	3I_OSS_WP E 41.71.013	41.50.013 41.52.013 41.53.100
3i Certain®	NP	41.14.014	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM 3I CER PRE-MILLED CoCr	3I_CER_NP E 41.71.014	41.50.014 41.52.014 41.53.100
	RP	41.14.015	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM 3I CER PRE-MILLED CoCr	3I_CER_RP E 41.71.015	41.50.015 41.52.015 41.53.100
	WP	41.14.016	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM 3I CER PRE-MILLED CoCr	3I_CER_WP E 41.71.016	41.50.016 41.52.016 41.53.100




	Platform	Analogue	Category	System	Kit	Scan Abutments
Straumann® Bone Level	NC	41.14.043	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM STR BLV PRE-MILLED CoCr	STR_BLV_NC E 41.71.043	41.50.043 41.52.043 41.53.100
	RC	41.14.044	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM STR BLV PRE-MILLED CoCr	STR_BLV_RC E 41.71.044	41.50.044 41.52.044 41.53.100
Zimmer Screw-Vent®	NP	41.14.017	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM ZIM SCV PRE-MILLED CoCr	ZIM_SCV_NP E 41.71.017	41.50.017 41.52.017 41.53.100
	RP	41.14.018	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM ZIM SCV PRE-MILLED CoCr	ZIM_SCV_RP E 41.71.018	41.50.018 41.52.018 41.53.100
	WP	41.14.019	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM ZIM SCV PRE-MILLED CoCr	ZIM_SCV_WP E 41.71.019	41.50.019 41.52.019 41.53.100
Astra tech implant system™ EV	EV/3.6	41.14.059	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM AST EVO PRE-MILLED CoCr	AST_EVO 3.6_E 41.71.059	41.50.059 41.52.059 41.53.100
	EV/4.2	41.14.060	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM AST EVO PRE-MILLED CoCr	AST_EVO 4.2_E 41.71.060	41.50.060 41.52.060 41.53.100
	EV/4.8	41.14.061	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM AST EVO PRE-MILLED CoCr	AST_EVO 4.8_E 41.71.061	41.50.061 41.52.061 41.53.100



Astra tech implant Osseospeed™	RP	41.14.024	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM AST OSS PRE-MILLED CoCr	AST_OSS_RP E 41.71.024	41.50.024	41.52.024	
							41.53.100	
	WP	41.14.025	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM AST OSS PRE-MILLED CoCr	AST_OSS_WP E 41.71.025	41.50.025	41.52.025	
							41.53.100	
Platform		Analogue	Category	System	Kit	Scan Abutments		
Friadent XIVE®	NP	41.14.038	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM DEN XIV PRE-MILLED CoCr	DEN_XIV_NP E 41.71.038	41.50.038	41.52.038	
							41.53.100	
	RP	41.14.039	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM DEN XIV PRE-MILLED CoCr	DEN_XIV_RP E 41.71.039	41.50.039	41.52.039	
						41.53.100		
	WP	41.14.040	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM DEN XIV PRE-MILLED CoCr	DEN_XIV_WP E 41.71.040	41.50.040	41.52.040	
							41.53.100	
Ankylos® C/X	All platforms, height GH 2,5	41.14.048	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM DEN ANK PRE-MILLED CoCr	DEN ANK E GH2 41.71.250	41.50.049	41.52.049	
		41.14.049				41.53.100		
Osstem® TS	Mini Yellow	41.14.072	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM OSS TS PRE-MILLED CoCr	OSS_TS MINI_E 41.71.072	41.50.072	41.52.072	
							41.53.100	
	STD Regular	41.14.073	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM OSS TS PRE-MILLED CoCr	OSS_TS STD_E 41.71.072	41.50.073	41.52.073	
							41.53.100	



	Platform	Analogue	Category	System	Kit	Scan Abutments
Camlog®	3.3	41.14.064	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM CAMLOG PRE-MILLED CoCr	CAM_3.3 E 41.71.064	41.52.064 41.53.100
	3.8	41.14.065	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM CAMLOG PRE-MILLED CoCr	CAM_3.8 E 41.71.065	41.50.065 41.52.065 41.53.100
	4.3	41.14.066	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM CAMLOG PRE-MILLED CoCr	CAM_4.3 E 41.71.066	41.50.066 41.52.066 41.53.100
	5.0	41.14.067	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM CAMLOG PRE-MILLED CoCr	CAM_5.0 E 41.71.067	41.52.067 41.53.100
Biohorizns® External	RP/4.2	41.14.021	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM BH EXT PRE-MILLED CoCr	BH_EXTERNAL_RP E 641.71.021	41.50.021 41.52.021 41.53.100
	WP/5.2	41.14.022	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM BH EXT PRE-MILLED CoCr	BH_EXTERNAL_WP E 41.71.022	41.50.022 41.52.022
MIS® Seven	SP	41.14.095	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM MIS SEV PRE-MILLED CoCr	MIS_SEV_SP_E 41.71.017	41.50.017 41.52.017 41.53.100
	WP	41.14.096	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM MIS SEV PRE-MILLED CoCr	MIS_SEV_WP_E 41.71.018	41.50.018 41.52.018 41.53.100
Megagen AntRidge®		41.14.057	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM MGN ANY PRE-MILLED CoCr	MGN ANY E 41.71.057	41.50.057 41.53.100
Dentium Superline™ & Implantium	RP	41.14.090	CENDRES METAUX PRE-MILLED ABUTMENT CoCr	CM DENTIUM PRE-MILLED CoCr	DENTIUM Ti base E 41.71.090	41.52.090 41.53.100



All CM Connect products are manufactured according to ISO 9001:2015 and ISO 13485:2016 with CE marking for all products Class IIb

All CM Connect AURUMBase® screws and screwdrivers are based on the standard ISO 10664

3i® and 3i Certain®, Miniplant®, MicroMiniplant™ and Osseotite® are registered trademarks of 3i Implant Innovations, USA
ASTRA TECH™, OsseoSpeed™, ANKYLOS® C/X, XiVE®, FRIADENT®, ASTRATECH implant system™ EV and UniAbutment®
are registered trademarks of DENTSPLY SIRONA Implants

BioHorizons® is a registered trademark of BioHorizons, INC.

Brånemark System®, NobelReplace® and Replace Select™, NobelActive™, NobelReplace® CC, NobelSpeedy™ and Multi-Unit are
registered trademarks of Nobel Biocare AB, Sweden

ITI®, Straumann® and synOcta® are registered trademarks of Straumann Holding AG, Switzerland

MIS® is a registered trademark of MIS Implant Technologies Ltd.

Zimmer Screw-Vent® and SwissPlus® are registered trademarks of Zimmer Dental INC, USA

Camlog® and Conelog® are registered trademarks of CAMLOG Biotechnologies AG

Megagen® is a registered trademark of MEGAGEN IMPLANT CO. LTD

TORX® is a trademark of ACUMENT INTELLECTUAL PROPERTIES, LLC

Bego Semados® is a registered trademark of BEGO GmbH & Co. KG

Anthogyr® is a registered trademark of Anthogyr SAS

Osstem® is a registered trademark of OSSTEM IMPLANT CO. LTD

Dentium® is a trademark of Dentium Co. Ltd.

Biotech Kontakt® is a trademark Biotech Dental SAS.

BTI® is a trademark BTI Biotechnology Institute.

CM Connect products can be used only by licensed dentists, physicians and labs.

Please refer ever to the instructions of use available on the website

www.cmsa.ch

or supplied with the product.

Some products may not be available in your country. Some of the products are cleared for US and Canada.

Please contact your local distributor for more details.